

CERTIFICATE OF COMPLIANCE

Certificate Number 20150601-E182560
Report Reference E182560-20150529
Issue Date 2015-JUNE-01

Issued to: Astec International Ltd - Philippine Branch
16TH FL, LU PLAZA, 2 WING YIP ST
KWUN TONG KOWLOON HONG KONG

**This is to certify that
representative samples of**

COMPONENT - POWER SUPPLIES, MEDICAL AND
DENTAL

Switching Power Supply, models 73-580-0001i and iMP8-
abbc-abbc-abbc-abbc-abbc-abbc-XX

Note:"abbc" is any alphanumeric character or blank for
specific model designation.

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety: ANSI/AAMI ES60601-1 & CAN/CSA-C22.2 No. 60601-1
Medical Electrical Equipment - Part 1: General
Requirements for Basic Safety and Essential Performance

Additional Information: See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's
Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance
capabilities and are intended for use as components of complete equipment submitted for investigation rather
than for direct separate installation in the field. The final acceptance of the component is dependent upon its
installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please
contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>

