



Ref. Certif. No.

SG PSB-MD-00100M2

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Switching power supply unit (Switching Power Supply)
Name and address of the applicant	Astec International Ltd. 16th Floor, Lu Plaza, 2 Wing Yip Street, Kwun Tong Kowloon, HONG KONG
Name and address of the manufacturer	Astec International Ltd. 16th Floor, Lu Plaza, 2 Wing Yip Street, Kwun Tong, Kowloon, HONG KONG
Name and address of the factory	Astec Power Philippines Inc., Main Avenue Corner Rd "J", Cavite Export Processing Zone, Rosario,, Cavite, PHILIPPINES For further information please see attachment
Ratings and principal characteristics	Rated Input : 395±5VDC, 8.5A Rated Output : See test report for details Construction : Built-in Protection Class : I Degree of Protection : IPX0
Model/type Ref.	73-936-0048, 73-936-0125, 73-936-0080, 73-936-0024, 73-936-0250, 73-936-0012, 73-936-0200, 73-936-0032
Additional information (if necessary)	Certificate SG PSB-MD-00100M1 issued on 2017-05-17 is replaced by this version due to technical changes
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012
as shown in the Test Report Ref. No. which forms part of this certificate	TÜV SÜD PSB Pte Ltd 211-600487-200

This CB Test Certificate is issued by the National Certification Body

Date, 2017-12-19
CBS 17 12 13890 02903

(Yager Bi)



TÜV SÜD PSB Pte Ltd · 1 Science Park Drive · Singapore 118221

Product Service

Additional factory information:

Name and address of the factory
Nom et adresse de l'usine


Astec Power Philippines Inc
104 Laguna Boulevard, Laguna Technopark, 4026
Sta Rosa Laguna, Philippines

Zhongshan General Carton Box Factory Co., Ltd.
No.62, Qi Guan Road West, Shiqi District, 528400,
Zhongshan City, Guangdong,
People's Republic of China

Astec Electronics (Luoding) Co. Ltd.
No. 68 Baocheng Road East, Fucheng Luoding,
527200 Guangdong,
People's Republic of China

Additional Information:

*The risk management requirements of the
standard were not addressed.*



Date, 2017-12-19
CBS 17 12 13890 02903

Signature : (Yager Bi)



Product Service

TÜV SÜD PSB Pte Ltd • 1 Science Park Drive • Singapore 118221