



Ref. Certif. No.

SG-MD-00218

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product
Produit

Switching power supply unit
(Open Frame Power Supply for Building-in)

Name and address of the applicant
Nom et adresse du demandeur

Astec International Limited - Philippine Branch
3rd & 4th Floor, Techno Plaza One Bldg, #18 Orchard Road, Eastwood
City Cyberpark, Bagumbayan
1110 Quezon City, PHILIPPINES

Name and address of the manufacturer
Nom et adresse du fabricant

Astec International Limited - Philippine Branch, 3rd & 4th Floor, Techno
Plaza One Bldg, #18 Orchard Road, Eastwood City Cyberpark,
Bagumbayan, 1110 Quezon City, PHILIPPINES

Name and address of the factory
Nom et adresse de l'usine

Astec Power Philippines Inc., Main Avenue Corner Rd "J", Cavite Export
Processing Zone, Rosario,, Cavite, PHILIPPINES
For further information please see attachment

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Rated DC Input : 100-250VAC, 50/60Hz, 1.6A or
140-300VDC, 1A
Rated DC Output : +5VDC, 5.0A; +15VDC, 2.5A;
-15VDC, 0.7A
Maximum output power: 40W with
convection cooling, 55W with 30CFM
forced-air cooling

Protection Class : I
Degree of Protection: IPX0

Trade mark (if any)
Marque de fabrique (si elle existe)

ASTEC

Model/type Ref.
Ref. de type

LPT45-M

Additional information (if necessary)
Information complémentaire (si nécessaire)

See Test Report for National Differences and Group Differences

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la


IEC 60601-1:2005

as shown in the Test Report Ref. No.
which forms part of this certificate
comme indiqué dans le Rapport d'essais numéro
de référence qui constitue une partie de ce
certificat

TÜV SÜD PSB Pte Ltd
211-2312052-000

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date, 2012-05-18
CBS 12 05 51485 01082


(Jimmy Huang)



TÜV SÜD PSB Pte Ltd · 1 Science Park Drive · Singapore 118221

PSB Singapore

Additional factory information:

Name and address of the factory
Nom et adresse de l'usine

Astec Power Philippines Inc.
104 Laguna Boulevard, Laguna Technopark, 4026
Sta Rosa Laguna
Philippines

Date, 2012-05-18
CBS 12 05 51485 01082

Signature :

Jimmy Huang